

CERTIFICATE NO.

PL-26-00059

VERIFICATION KEY

NWVZPUY4

Scan or enter both at
prymalab.net/testing/verify

CERTIFICATE OF ANALYSIS · LOT RELEASE · ISSUED FOR BLUTIDES

KPV 10 mg

Lot PL-2609-059 · Lyophilized powder · Label claim 10 mg/vial · Distributed as KPV

TESTED IN-HOUSE BY PRYMALAB

In-house analytical result,
not an independent third-party test

LOT / BATCH NUMBER	DATE MADE	DATE OF ANALYSIS	DATE ISSUED
PL-2609-059	2026-09-15	2026-09-16	2026-10-03
RETEST / EXPIRY	UNITS IN LOT / SAMPLED	STORAGE	CERTIFICATE TYPE
2028-09-15	250 / 2	-20 °C, desiccated, dark	Lot COA, 2 pages

COMPOUND	CAS NUMBER	MOLECULAR FORMULA	THEORETICAL MASS (AVG.)
KPV	67727-97-3	C ₁₆ H ₃₀ N ₄ O ₄	342.4 Da
Lys-Pro-Val; α-MSH(11-13)			
SEQUENCE			
H-KPV-OH (3 residues; free N- and C-termini)			

PURITY · RP-HPLC, AREA %

99.79%

Specification 99.0 % • main peak at 3.31 min

IDENTITY · OBSERVED MASS

342.6 Da

Theoretical 342.4 Da • +0.2 Da (584 ppm)

OVERALL CONCLUSION


PASS

Meets every specification for the tests performed

TEST	METHOD	SPECIFICATION	RESULT	CONCLUSION
Appearance	Visual inspection	White to off-white lyophilized powder	White lyophilized powder	PASS
Identity (mass)	LC-MS, ESI positive	Theoretical ± 1.0 Da	342.6 Da	PASS
Purity	RP-HPLC, UV 214 nm, area%	99.0%	99.79%	PASS
Largest impurity	RP-HPLC, area%	1.0%	0.09% at 4.14 min	PASS
Water content	Karl Fischer	5.0%	1.7%	PASS
Bacterial endotoxin	LAL, kinetic	≤ 5.0 EU/mg	0.05 EU/mg	PASS
Sterility / bioburden	USP <71>, membrane filtration	No growth after 14 days	No growth; conforms	PASS

1 HPLC CHROMATOGRAM

Data file: KPV_PL-2609-059_20260916_01 · Column C18, 150 × 4.6 mm · UV 214 nm



2 HPLC METHOD · KPV_PL-2609-059_20260916_01

INSTRUMENT	DATA SYSTEM	DETECTOR
Waters Alliance e2695, SN E2695B742	Empower 3, Build 3471	Waters 2998 PDA
DETECTION WAVELENGTH	COLUMN	COLUMN PART / SERIAL
214 nm	C18, 150 × 4.6 mm, 3.5 µm	186003326 / USHX18211
MOBILE PHASE A	MOBILE PHASE B	FLOW RATE
0.1% TFA in water	0.1% TFA in MeCN	1.0 mL/min
COLUMN TEMPERATURE	INJECTION VOLUME	RUN TIME
30 °C	20 µL	10 min
SAMPLE DILUENT	SAMPLE CONCENTRATION	BLANK INJECTED
Water, 0.1% TFA	0.5 mg/mL	Yes
REFERENCE STANDARD		
In-house KPV reference, lot PL-RS-KPV-01, RT 3.29 min		

3 GRADIENT PROGRAM

TIME (MIN)	% A	% B
0.0	95	5
3.0	95	5
7.0	55	45
8.0	20	80
9.0	20	80
10.0	95	5

4 PEAK TABLE

PEAK	RT (MIN)	AREA	AREA %
1	2.49	2,975	0.07
2	3.31	4,241,075	99.79
3	4.14	3,825	0.09
4	4.97	2,125	0.05
Total		4,250,000	100.00

Threshold 0.05% area · Blank subtracted: Y · Purity = main peak area ÷ total integrated area × 100, to two decimals; main peak shaded.

5 IDENTITY BY MASS SPECTROMETRY

INSTRUMENT	IONIZATION / MODE	OBSERVED M/Z (CHARGE)
Waters Acquity SQD	ESI, positive	343.3 (1+ [M+H]) · 365.3 (1+ [M+Na])
DECONVOLUTED MASS	THEORETICAL MASS	DIFFERENCE
342.6 Da	342.4 Da	+0.2 Da (584 ppm)
DATA FILE	ACCEPTANCE	
KPV_PL-2609-059_MS_20260916_01	Within ± 1.0 Da	

6 ADDITIONAL TESTS

TEST	PERFORMED BY	DATE	REPORT / FILE
Water content	PrymaLab	2026-09-17	RPT-KF-PL-2609-059
Endotoxin	PrymaLab	2026-09-18	RPT-ET-PL-2609-059
Sterility	PrymaLab	2026-09-30	RPT-ST-PL-2609-059

A test with no row here was not performed on this lot. Outside-laboratory reports, when used, are attached and named below.

7 REVIEW AND APPROVAL

D. Reynolds	S. Calloway	M. Phelps
Analyst · performed the testing	Reviewer · checked against raw data	Lab Director · approved for release
Date: 2026-09-16	Date: 2026-10-02	Date: 2026-10-03

Raw data retained at: Empower 3 (Build 3471), project PL-2609, archive PL-ARC-2026; 7 years. Outside laboratory reports attached: None

This certificate applies only to the lot named above and describes the units sampled from it; it is not a guarantee for any other lot. Results were obtained by PrymaLab LLC's own laboratory (PrymaLab Analytical) and are not an independent third-party test unless an outside report is attached. This copy is issued for BluTides, a distributor of the lot; the brand shown does not change the laboratory, the method or the results. Reference values (CAS, formula, theoretical mass, sequence) are literature values for the named compound. Authenticity: scan the QR code or enter certificate number PL-26-00059 and key NWVZPUY4 at prymalab.net/testing/verify; any copy that does not match the record shown there is not a PrymaLab certificate. PrymaLab LLC products are supplied for laboratory research use only and not for human or veterinary use.